



Research Article

A Rapid and High-Sensitive Real-Time Reverse Transcription-Polymerase Chain Reaction Assay Used for the Detection of Severe Acute Respiratory Syndrome Coronavirus 2

Xiaomin Chen^{1,2}, Yan Xu¹, Jing Tian¹, Xueling Li¹, Hui Liang¹, Ang Gao^{1,2},
Zexi Liu¹, Dicheng Yang¹, Qi Shen¹, Daxiang Cui^{1,2}✉

¹National Engineering Research Center for Nanotechnology, Shanghai, China.

²Institute of Nano Biomedicine and Engineering, Shanghai Engineering Research Centre for Intelligent Diagnosis and Treatment Instrument, Department of Instrument Science and Engineering, School of Electronic Information and Electrical Engineering, Shanghai Jiao Tong University, Shanghai, China.

✉ Corresponding author. E-mail: dx cui@s jtu.edu.cn

Received: Aug. 24, 2020; **Accepted:** Oct. 27, 2020; **Published:** Oct. 27, 2020

Citation: Xiaomin Chen, Yan Xu, Jing Tian, Xueling Li, Hui Liang, Ang Gao, Zexi Liu, Dicheng Yang, Qi Shen, and Daxiang Cui, A Rapid and High-Sensitive Real-Time Reverse Transcription-Polymerase Chain Reaction Assay Used for the Detection of Severe Acute Respiratory Syndrome Coronavirus 2. *Nano Biomed. Eng.*, 2020, 12(4): 311-315.

DOI: 10.5101/nbe.v12i4.p311-315.

Abstract

The coronavirus disease 2019 (COVID-19) pandemic caused by severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) is a public health emergency of international concern. Real-time reverse transcription-polymerase chain reaction (RT-PCR) is widely used as the gold standard method for the diagnosis of SARS-CoV-2 infection. However, the reliability of current real-time RT-PCR assays is questioned due to some false-negative reports. In this study, we improved the real-time RT-PCR method based on three target regions (ORF1ab, E, and N) of SARS-CoV-2. Results showed that real-time RT-PCR assays herein could complete detection within one hour after viral RNA preparation and had high sensitivity down to 5 copies of viral RNA. In addition, six clinical specimens were detected to evaluate the availability of this method. Among them, four samples were 3-plex SARS-CoV-2 positive and two were negative by real-time RT-PCR. The sensitivity was 100% (4/4), and specificity was 100% (2/2). These results demonstrate that we develop a rapid and high-sensitive real-time RT-PCR method for SARS-CoV-2 detection, which will be a powerful tool for COVID-19 identification and for monitoring suspected patients.

Keywords: COVID-19, SARS-CoV-2, RT-PCR, Diagnostic, Coronavirus

Introduction

The coronavirus disease 2019 (COVID-19), caused by severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2), has spread around the world and has become a public health emergency of international concern [1, 2]. To date (20 August 2020), up to

22,213,869 COVID-19 cases have been confirmed, including more than 781,677 deaths [3]. In the absence of vaccines and antivirals, early diagnosis and quarantine are effective measures to cure patients and control the spread of COVID-19. The viral nucleic acid detection by real-time reverse transcription-polymerase chain reaction (RT-PCR) based on the

TaqMan Probe assay is widely employed as the gold standard method for the diagnosis of SARS-CoV-2 infection [4, 5]. Advantages of real-time RT-PCR assays are high-throughput and sensitive, and even to quantify viral loads in patients based on a standard curve for monitoring disease progression [6]. Since the global outbreak of COVID-19, real-time RT-PCR assays based on various prime/probe sets are quickly developed for SARS-CoV-2 detection [7-10]. However, their reliability is questioned due to the presence of false-negative results in some patients and positive results in some confirmed cases after recovery [11-14]. Thus, it is very important to further optimize and improve current RT-PCR assays for better screening suspected patients and preventing COVID-19 spread.

SARS-CoV-2 is an enveloped positive-sense single-stranded RNA (+ ssRNA) virus [1, 7]. Phylogenetic analyses reveal that SARS-CoV-2 belongs to β -coronavirus and its nucleotide sequence is close to bat-derived SARS-like coronaviruses [15, 16]. In order to distinguish from other high-pathogenic viruses, avoid viral mutation, and even detect the low viral load, SARS-CoV-2 is usually identified by multiple RT-PCR targeting different viral regions [4]. In this study, we optimized the real-time RT-PCR assay using four primer/probe sets targeting the open reading frame 1ab (ORF1ab), envelope protein (E), and nucleocapsid protein (N) regions of SARS-CoV-2 to improve COVID-19 detection. We also evaluated the sensitivity of this method. In addition, six clinical specimens were analyzed for the availability of this method.

Experimental

Preparation of the artificial SARS-CoV-2 gene

The synthesized sequence of the *ORF1ab*, *E*, and *N* gene of SARS-CoV-2 (BIOLIGO, Shanghai, China) were used to optimize real-time RT-PCR test. The SARS-CoV-2 RNA transcribed *in vitro* Reference Material (Shanghai Institute of Measurement and Testing Technology, Shanghai, China) containing *ORF1ab* (13321-15540, GenBank No. MT027064.1), the full-length *E* gene and *N* gene of SARS-CoV-2 quantitated by digital PCR were used to evaluate the sensitivity of real-time RT-PCR.

RNA extraction

Total RNA was extracted from the inactivated clinical specimens from Shanghai Center Clinical Laboratory (Shanghai, China) using the TaKaRa MiniBEST Viral RNA/DNA Extraction Kit Ver.5.0

(Takara, Japan) following manufacturer's instructions. The extracted RNA was stored at -80 °C.

Primers and probes

Four high-specific primer/probe sets targeting the ORF1ab, E, and N regions of SARS-CoV-2 issued by China Center for Disease Control and Prevention (CDC) [7] and Corman et al. [8] were synthesized by BIOLIGO (Shanghai, China). Target 1 (*ORF1ab*), forward: 5'-CCCTGTGGG TTTTACACTTAA-3', reverse: 5'-ACGATTG TGCATCAGCTGA-3', probe: 5'-FAM-CC GTCTGCGGTATGTGGAAAGGTTATGG-BQ1-3'; Target 2 (*E*), forward: 5'-ACAGGTACGTTAATAGTT AATAGCGT-3', reverse: 5'-ATATTGCAGCAGTACG CACACA-3', probe: 5'-FAM-ACACTAGCCATC CTTACTGCGCTTCG-BQ1-3'; Target 3 (*N1*), forward: 5'-GGGGAAGTCTCCTGCTAGAAAT-3', reverse: 5'-CAGACATTTTGCTCTCAAGCTG-3', probe: 5'-FAM-TTGCTGCTGCTTGACAGATT-BQ1-3'; Target 4 (*N2*), forward: 5'-CACATTG GCACCCGCAATC-3', reverse: 5'-GAGGAACGA GAAGAGGCTTG-3', probe: 5'-FAM-ACTTCCT CAAGGAACAACATTGCCA-BQ1-3'.

Real-time RT-PCR assay

Real-time RT-PCR assay was developed using the One Step PrimerScript™ RT-PCR kit (Takara, Japan). A 25 μ L reaction mixture contained 12.5 μ L of 2X PCR Buffer, 0.5 μ L of reverse transcriptase, 0.5 μ L of *Taq* DNA Polymerase, 200 nM of probes, 1 μ L of primers, and 5 μ L of RNA. Real-time RT-PCR was performed using Stratagene Mx3000P (Agilent, Palo Alto CA, USA). The optimal reaction procedure was 42 °C for 5 min, 95 °C for 10 s, followed by 40 cycles of 95 °C for 5 s and 60 °C for 30 s.

Statistical analysis

Data were presented as the mean \pm standard deviation (SD). The Student's t-test was used in quantitative data analysis, and $P < 0.05$ was considered to indicate a statistically significant difference. All statistical analyses were performed with SPSS 20.0 software (SPSS, Inc., Chicago IL, USA).

Results and Discussion

Optimizing the real-time RT-PCR assay

Many commercial nucleic acid detection kits based on real-time RT-PCR assays have been widely used for SARS-CoV-2 detection, but there are differences

in their detection capability of weak positive samples [17, 18]. Currently, time-consuming and insensitive methods are not able to meet the requirement of large-scale molecular diagnosis of suspected patients and asymptomatic patients. Thus, to shorten the viral detection, we first performed real-time RT-PCR assays at three different thermal cycling time. As shown in Fig. 1, the highest amplification of *ORF1ab*, *E*, and *N2* at 400 nM of primers occurred in 35 s per thermal cycle, and *N1* occurred in 25 s. Due to a multiple test, the optimal time per thermal cycle was determined at 35 s, meaning the whole real-time RT-PCR assay spent one hour completing detection.

In order to increase the amplification products by real-time RT-PCR, we performed the reaction at different primer concentration from 200 nM to 400 nM. As presented in Fig. 2, the highest amplification of *ORF1ab* occurred at 400 nM of primers, and the other three genes had no significant differences in the range of 200 to 400 nM. Therefore, the optimal primer concentration of *ORF1ab* was 400 nM, and others adopted in the subsequent studies were 200 nM.




Fig. 1 Effect of thermal cycling time on real-time RT-PCR assays. Concentration of primers was 400 nM. * $P < 0.05$, ** $P < 0.01$, *** $P < 0.001$.




Fig. 2 Effect of primer concentration on real-time RT-PCR assays. * $P < 0.05$.

Sensitivity of the real-time RT-PCR assay

The performance of the aforementioned real-time RT-PCR assay was evaluated using ten-fold serial dilutions of the viral RNA ranging from 5×10^4 copies to 5 copies. As illustrated in Fig. 3, the cycle threshold (Ct) of real-time RT-PCR was linearly linked to the number of observed viral RNA, and regression analysis produced a determination coefficient (R^2) of 0.941~0.992. The World Health Organization (WHO) posts various high-specific primer/probe sets for SARS-CoV-2 detection developed at different institutions [4]. However, assays among these primer/probe sets have different ability to detect SARS-CoV-2 [19, 20]. Chantal et al. [19] found that the most sensitive primer/probe sets in the same RT-PCR reagents and conditions are Corman E (Charité-Universitätsmedizin Berlin Institute of Virology), HKU-ORF1 (Hong Kong University), HKU-N (Hong Kong University), CCDC-N (China CDC), 2019-nCoV_N1 (United States CDC), and 2019-nCoV_N3 (US CDC), which could partially detect SARS-CoV-2 at 1 (25%) and 10 (25-50%) copies of viral RNA. Different from reaction conditions described by Chantal et al. [19], four primer/probe sets in this study also represented different amplification efficiency. The reportable range of *ORF1ab* was $5 \sim 5 \times 10^4$ copies (Fig. 3(a)), and the limit of detection (LoD) of real-time RT-PCR using *ORF1ab* analyzed with twelve replicates was 5 copies of viral RNA (12/12) (Fig. 4(a)). The *E* assay also exhibited high sensitivity (Fig. 3(b)), but it was less sensitive than *ORF1ab* (Fig. 4(b)). In fact, LoD of real-time RT-PCR using *E* primer/probe set was more than 5 copies (5/12). Compared to the *ORF1ab* and *E* assays, the *N* assay was not reliable at low SARS-CoV-2 amounts. Among two N gene assays, the *N1* primer/probe set was more sensitive than *N2* set (Fig. 3(c) and 3(d)). To further confirm the improved real-time RT-PCR method, we utilized three primer/probe (*ORF1ab*, *E* and *N1*) sets in the subsequent studies.

Clinical sample detection

Six clinical specimens were detected using the optimal real-time RT-PCR assay based on *ORF1ab*, *E*, and *N1* primer/probe sets. Among them, four 3-plex positive samples were confirmed to be infected with SARS-CoV-2, while two were negative (Table 1). The sensitivity was 100% (4/4), and specificity was 100% (2/2). In addition, the results also showed that assays using *ORF1ab* or *E* primer/probe set performed better




Fig. 3 Sensitivity analysis of real-time RT-PCR. The reportable range of real-time RT-PCR using (a) *ORF1ab*, (b) *E*, (c) *NI*, and (d) *N2* primer/probe sets were determined. Data were representative of three independent experiments.




Fig. 4 (a) The limit of detection (LoD) of real-time RT-PCR based on *ORF1ab* primer/probe set. (b) Analysis of real-time RT-PCR using *ORF1ab* and *E* primer/probe sets to detect 9 copies of viral RNA. Data were analyzed with twelve replicates. ** $P < 0.01$.

Table 1 Clinical samples detected by real-time RT-PCR

Sample	Ct Value			Results
	<i>ORF1ab</i>	<i>E</i>	<i>NI</i>	
1	-	-	-	-
2	30.20	31.71	35.60	+
3	-	-	-	-
4	28.87	29.94	33.21	+
5	28.16	28.79	32.87	+
6	35.31	33.95	38.84	+

-: Negative; +: Positive.

than those using *NI* set (Table 1). Similar to most commercially available and laboratory-developed assays, we recommend that a 3-plex or 2-plex positive outcome is defined as SARS-CoV-2 positive, and not detected test result is negative. As for single-plex result, it is regarded as inconclusive and subsequently retested for the confirmation. Of course, a large scale of clinical study is necessary to prove the accuracy of

this improved real-time RT-PCR method.

Conclusions

In this study, we have developed a rapid and high-sensitive real-time RT-PCR method based on the *ORF1ab* (China CDC), *NI* (China CDC), and Corman *E* primer/probe sets to improve the current SARS-CoV-2 detection. This real-time RT-PCR assay can rapidly complete detection within one hour after viral RNA preparation. Moreover, the optimal multiplexed real-time RT-PCR assay exhibits high sensitivity down to 5 copies of viral RNA. In addition, our findings indicate that *ORF1ab* and *E* assays have a higher analytical sensitivity than *N* assays, which means that *N* assays are not a reliable confirmatory assay at low SARS-CoV-2 amounts.

Acknowledgements

This work was supported by Key Basic Research Program of China (No.2017YFA0205304), Nature Scientific Foundation of China (No.81602184), and Medical Engineering Cross Project of Shanghai Jiao Tong University (YG2017ZD12). This work was also supported by “the Belt and Road” young scientist exchange program of the Science and Technology Commission of Shanghai (Grant No. 18410741600).

Conflict of Interests

The authors declare that no completing interest exists.

References

- [1] R. Lu, X. Zhao, J. Li, et al., Genomic characterisation and epidemiology of 2019 novel coronavirus: implications for virus origins and receptor binding. *The Lancet*, 2020, 395(10224): 565-574.
- [2] L. Chen, W. Liu, Q. Zhang, et al., RNA based mNGS approach identifies a novel human coronavirus from two individual pneumonia cases in 2019 Wuhan outbreak. *Emerg Microbes Infect*, 2020, 9(1): 313-319.
- [3] World Health Organization, WHO Coronavirus Disease (COVID-19) Dashboard. 2020, <https://www.covid19.who.int/>.
- [4] World Health Organization, Molecular assays to diagnose COVID-19: Summary table of available protocols. *COVID-19: Laboratory and diagnosis. bioRxiv*, 2020, <https://www.who.int/>.
- [5] R. Liu, H. Han, F. Liu, et al., Positive rate of RT-PCR detection of SARS-CoV-2 infection in 4880 cases from one hospital in Wuhan, China, from Jan to Feb 2020. *Clin Chim Acta*, 2020, 505: 172-175.
- [6] F. Yu, L. Yan, N. Wang, et al., Quantitative detection and viral load analysis of SARS-CoV-2 in infected patients. *Clin Infect Dis*, 2020, 71(15): 793-798.
- [7] N. Zhu, D. Zhang, W. Wang, et al., A novel coronavirus from patients with pneumonia in China, 2019. *N Engl J Med*, 2020, 382(8): 727-733.
- [8] V.M. Corman, O. Landt, M. Kaiser, et al., Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-PCR. *Euro Surveill*, 2020, 25(3): 2000045.
- [9] B. Udugama, P. Kadhiresan, H.N. Kozlowski, et al., Diagnosing COVID-19: The Disease and Tools for Detection. *ACS Nano*, 2020, 14(4): 3822-3835.
- [10] U. Pandey, A.L. Greninger, G.R. Levin, et al., Improved molecular diagnosis of COVID-19 by the novel, highly sensitive and specific 2019-nCoV-RT-PCR real-time reverse transcription-polymerase chain reaction assay validated in vitro and with clinical specimens. *J Clin Microbiol*, 2020, 58(5): e00310-e00320.
- [11] C. Xie, L. Jiang, G. Huang, et al., Comparison of different samples for 2019 novel coronavirus detection by nucleic acid amplification tests. *Int J Infect Dis*, 2020, 93: 264-267.
- [12] T. Suo, X. Liu, J. Feng, et al., ddPCR: a more accurate tool for SARS-CoV-2 detection in low viral load specimens. *Emerg Microbes Infect*, 2020, 9(1): 1259-1268.
- [13] T. Ai, Z. Yang, H. Hou, et al., Correlation of Chest CT and RT-PCR Testing for Coronavirus Disease 2019 (COVID-19) in China: A Report of 1014 Cases. *Radiology*, 2020, 296(2): E32-E40.
- [14] P. Huang, T. Liu, L. Huang, et al., Use of Chest CT in combination with negative RT-PCR Assay for the 2019 novel coronavirus but high clinical suspicion. *Radiology*, 2020, 295(1): 22-23.
- [15] F. Wu, S. Zhao, B. Yu, et al., Complete genome characterisation of a novel coronavirus associated with severe human respiratory disease in Wuhan, China. 2020.
- [16] P. Zhou, X.L. Yang, X.G. Wang, et al., A pneumonia outbreak associated with a new coronavirus of probable bat origin. *Nature*, 2020, 579(7798): 270-273.
- [17] Y. Li, D. Li, B. Wang, et al., Comparison of detection performance between six novel coronavirus nucleic acid detection reagents. *Shandong Medical Journal*, 2020, 60(15): 14-17.
- [18] D. Xiong, L. Kan, M. Wang, et al., Evaluation of the consistency and detection capability of seven domestic 2019-nCoV nucleic acid detection kits. *Chin J Lab Med*, 2020, 43(8): 787-793.
- [19] B.F.V. Chantal, F.B. Anderson, A.W. Anne, et al., Analytical sensitivity and efficiency comparisons of SARS-CoV-2 RT-qPCR primer-probe sets. *Nat Microbiol*, 2020, 5(10): 1299-1305.
- [20] Y. Jung, G.S. Park, J.H. Moon, et al. Comparative Analysis of Primer-Probe Sets for RT-qPCR of COVID-19 Causative Virus (SARS-CoV-2). *ACS Infect Dis*, 6(9): 2513-2523.

Copyright© Xiaomin Chen, Yan Xu, Jing Tian, Xueling Li, Hui Liang, Ang Gao, Zexi Liu, Dicheng Yang, Qi Shen, and Daxiang Cui. This is an open-access article distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.